DOCUMENTED RADIATION SURVEY
UNIVERSITY OF MISSOURI-COLUMBIA

List all locations where the survey data exceed the levels listed below, if none indicate "None".
When laboratory ALARA or Survey Procedure action levels are exceeded, the actions taken must be documented in the comment section below

Instrument: Efficiency: (DPM=CPM/Eff) Calibration date:	Instrument: Calibration date:
Removable Contamination locations; Swipe(s) measurements greater than 200 dpm/100 cm^2:	Surface Contamination, locations with readings greater than 0.3 mR/hr @ 1 cm:
Comments:	Comments:

AUTHORIZED USER: _____ ROOM(S) & BLDG._____ Date: _____ Performed By:_____